

Defibrillator Post-use Checklist

ZOLL AED Plus
ZOLL AED Plus Serial Number: Defibrillator Coordinator: Defibrillator Location: <i>Complete the checklist and forward to the Occupational Health Team for retention.</i>

Refer to <i>ZOLL AED Plus Administrator's Guide</i> pages 4, 11, and 17-22 for detailed instructions. Contact the Occupational Health Team for assistance	Completed	Date
1. Unit and accessories are free from damage, dirt and contamination Clean and/or replace if necessary (refer to page 22)	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
2. New CPR-D-padz electrodes are connected or replacements have been ordered* (refer to page 11) New CPR-D-padz electrodes have passed expiration date	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
3. New batteries have been in installed or replacements have been ordered* (refer to pages 18-21) Defibrillator has passed <i>Battery Installation Self Test</i> . Batteries have not passed expiration date <i>Please note: All 10 batteries must be replaced at once. Do not press the Battery Reset Button if not all the batteries are new). Affix a sticker with batteries' expiry date next to the ON/OFF button</i>	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
4. Manual self-test conducted (refer to page 17) <i>Hold ON/OFF button down for at least 5 seconds to start a manual self-test</i> Unit will report status of the self-test	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
5. Status Indicator shows an illuminated green check mark <input checked="" type="checkbox"/> within 4-5 seconds of the unit being turned on and off (refer to page 4)	<input type="checkbox"/> YES <input type="checkbox"/> NO/...../.....
Signature: <i>(Print name if different to Defibrillator Coordinator listed above)</i>/...../.....	

* Order replacement electrode pads immediately post use. While awaiting pad replacement, place 'OUT OF ORDER' signage on the defibrillator.